

PDVS2mini - TEST / CALIBRATION RECORD

FUNCTION	COUNTS/SETTING	VOLTAGE OUTPUT
0.00000	2949	-00.000008865
1.00000	102076	01.000008663
2.00000	201200	01.999999274
3.00000	300326	02.999997945
4.00000	399453	03.999998928
5.00000	498580	05.000001886
6.00000	597708	06.000005751
7.00000	696836	07.000008943
8.00000	795960	07.999995223
9.00000	895086	09.000001502
10.00000	994209	10.000004270
Out.Vdc	1076	
Batt.Vdc	2790	
Charge mA	1178	
Battery Low/Shutdown	13.0/12.5Vdc	
Charge Enable	YES	
Charge Overload mA	400mA	
Charge Full mA	70mA	
Serial Comms	OK	
Soak Test Duration	168hrs	
External DC Input	OK	
Battery Charge/Discharge Test	OK	
Serial Number	0163	
Ambient Temperature degC	22.5	
Date	21 Feb. 2021	

Cal = HP3458A S/N=2823A24074

GPIB/Visa/PDVS2mini App V2.3